

January 5, 1961

To: Members of IRE Subcommittee 27.4

Enclosed please find a draft of Section 5 and Section 6 of our proposed Standard of Measuring Interference Output. These incorporate the points discussed at the December 2, 1960 meeting.

Please review these before our next meeting.

Very truly yours,

V.J. Mancino

V.J. Mancino, Chairman
Subcommittee, 27.4

VJM:js

JMAN PRO

1913 12 15

Dear Sir,

I have the pleasure to acknowledge the receipt of your letter of the 12th inst. in relation to the above mentioned matter.

The same has been referred to the proper authorities for their consideration.

I am, Sir, very respectfully,
Yours truly,
J. M. Anderson

3/10/61

December 20, 1960

IRE 27.4 Proposed Standards

Methods of Measurement of Transmitter External Wiring Spurious Output5. Spurious Output appearing on Transmitter External Wiring

5.1 General. The Spurious Output appearing on the external wiring of a transmitter which may consist of harmonic (or non-harmonic) components, may be measured by measuring the voltage developed across a known impedance network with the use of a calibrated frequency-selective voltmeter or receiver.

5.2 Equipment required:

5.2.1 Impedance Network. The spurious output appearing on external wiring should be measured with the use of an impedance network whose characteristic is known. This network shall be inserted in the external ~~wire~~ ^{wire} whose spurious output is to be measured, at a convenient point close to the point of entry of the wire to the transmitter hardware, such that the spurious output voltage is developed from the wire to ground. The network must have a connection means for cables to the calibrated frequency-selective voltmeter or receiver. The circuit of this network is shown in figure 5A ^{and its impedance curve is shown in Figure 5B.} (Make same as Pg 20 of Ref. I - 26600)

5.2.2 Variable Attenuator. A calibrated ^{variable} attenuator is needed to provide a means for adjusting the output indication level of the measuring instrument so as not to overload the frequency selective, voltmeter or receiver. *It may also be useful in checking for spurious responses*

5.2.3 Frequency-selective voltmeter or receiver. A frequency-selective voltmeter or receiver (may be more than one) which can tune to the carrier and any spurious output frequency of interest, is needed.

If a receiver is used, it must have an output indicator. The receiver cabinet should incorporate good shielding techniques and the power lines should be well filtered. ✓

5.2.4 Coaxial Switches. Coaxial switches (or suitable means for changing connections) may be required when using^a calibrated signal generator for calibration purposes. ✓

5.2.5 Measuring Equipment Enclosure. To prevent pickup of extraneous radiations during the measurements, the measuring equipment should (if necessary) be enclosed within a suitable shielded enclosure and the signal from the pickup device brought into the shielded enclosure through a well-shielded cable. ✓

5.2.6 Calibrated Signal Generator. A calibrated signal generator (or generators) to cover the carrier frequency and any spurious frequencies of interest is needed. ✓

5.3 Measurement Procedure.

5.3.1 Standard Method. NOTE: The standard method measures the voltage of the spurious output developed across a known impedance network. *Then*

5.3.1.1 Connect the equipment as shown in figure 3.

5.3.1.2 Operate the transmitter under test in its intended manner with its output connected to a shielded dummy load.

5.3.1.3 Tune the frequency-selective voltmeter or receiver through the frequency range of interest with the^{calibrated} variable attenuator adjusted for maximum sensitivity of the measuring circuit. When a spurious output is found, adjust the attenuator to obtain a suitable reference indication on the output indicator of the measuring instrument.

CAUTION: In all frequency-selective voltmeters or receivers spurious

responses may occur by: (1) Desensitization of the receiver by the entry of a strong off-channel signal through the antenna input, (2) ~~the~~ the entry of a strong on-channel signal through the receiver case or power lines, ~~and~~ ^{such signals} by-passing ~~the~~ ^{the} calibrated input attenuator. These responses must be known or determined for the particular device used.

In addition care must be taken to insure that the spurious signal being measured can be actually attributed to the equipment under test. This is easily determined by momentarily turning off the equipment under test.

5.3.1.4 If a frequency-selective voltmeter is used, calibrate it according to the recommended manufacturer's procedure and measure the spurious output signal. If a receiver is used it must be calibrated by means of appropriate calibrated signal generators.

5.3.1.5 The level of the spurious output is calculated as follows:

5.3.1.5.1 CW Spurious Output. CW Spurious Output Level

(DB above 1uV) = meter reading in DB above 1uV (or substituted signal generator reading) + cable loss in DB (*between the impedance network and the calibrated variable attenuator*)

5.3.1.5.2 Broadband Spurious Output

Broadband Spurious Output Level (DB above 1uV per MC) = meter reading in DB above 1uV (or substituted signal generator reading) + Cable loss (DB) - $10 \log_{10}$ (impulse bandwidth in MC).

5.3.2 Alternate Method. NOTE: The alternate method is used when

it is inconvenient to use the standard method due to reasons of inaccessibility or ~~due to the large current carrying capacity of the external wiring~~ ^{other considerations}. For this method a clamp-on current ~~probe~~ ^{transformer} of known

characteristics is used. The spurious output is determined by measuring the current induced in the ^{clamp-on transformer} ~~current~~. Due to the inherent

limitations of the method the results will require interpretation.

5.3.2.1 Connect the equipment as shown in figure 4.

5.3.2.2 Operate the transmitter under test in its intended manner with its output connected to a ^{shielded} dummy load.

5.3.2.3 Tune the frequency-selective voltmeter or receiver through the frequency range of interest with the ^{calibrated} variable attenuator adjusted for maximum sensitivity of the measuring circuit. When a spurious output is found, adjust the attenuator to obtain a suitable reference ^{indication} on the output indicator *of the measuring instrument.*

CAUTION: In all frequency-selective voltmeters or receivers spurious responses may occur by: (1) Desensitization of the receiver by the entry of a strong off-channel signal through the antenna input. (2) ~~the~~ the entry of a strong on-channel signal through the receiver case or power lines. ^{such signals} ~~by-passing~~ ^{the} calibrated input attenuator. These responses must be known or determined for the particular device used.

In addition care must be taken to insure that the spurious signal being measured can be actually attributed to the equipment under test. This is easily determined by momentarily turning off the equipment under test.

5.3.2.4 If a frequency-selective voltmeter is used, calibrate it according to the recommended manufacturer's procedure and measure the spurious output signal. If a receiver is used it must be calibrated by means of appropriate calibrated signal generators.

5.3.2.5 The level of the spurious output is calculated as follows:

5.3.2.5.1 CW Spurious Output. Spurious Output Level (DB above μA) =
 Meter Reading in DB above μV (or substituted signal generator reading)
 + cable loss in DB ^(between the clamp-on current transformer and the calibrated variable attenuator)
 + ^{current} voltage to ^{voltage} current conversion factor in DB.

5.3.2.5.2 Broadband Spurious Output. Spurious Output Level

(DB above 1uA per MC) = Meter reading in DB above 1uV (or substituted
signal generator reading) + cable loss in DB + ~~voltage~~^{current} to ~~current~~^{voltage} conversion
factor in DB - $10 \log_{10}$ (impulse bandwidth in MC)




Figure 3




Figure 4

3/10/61

IRE 27.4 Proposed Standards

Methods of Measurement of Transmitter Cabinet Radiation6. Spurious Output Due to Transmitter Cabinet Radiation.

6.1 General. The Spurious Output radiated from a transmitter, which may consist of harmonic (or non-harmonic) components, may be measured by measuring the voltage induced in a receiving antenna with the use of a calibrated frequency-selective voltmeter or receiver.

6.2 Equipment Required:

6.2.1 Pickup Device. The radiated spurious output should be measured with a pickup device which is connected to a calibrated frequency-selective voltmeter or receiver by an impedance matching network or balun. This pickup device may be a rod or loop for frequencies below 25MC, a resonant dipole for frequencies from 25MC to 1000MC, and a horn antenna for frequencies above 1000MC. The pickup device must be calibrated over its useful frequency range so that it may be used to ~~read open circuit voltage~~ *measure the intensity of the electromagnetic field.* (antenna induced voltage).

6.2.2 Interference-Free Area. The radiated spurious output shall be measured in an area sufficiently free from ambient interference and physical obstructions for the purposes of this measurement. It is desirable that the ambient interference level during testing be at least 6 db below the interference limits specified in the appropriate specifications. However, in the event that at the time of measurement the levels of ambient interference plus the spurious output of the item under test are not above the specified limit, such tested item shall be considered to comply with the specified requirements. In addition any frequency whose identity is definitely established such as a broadcast station is exempt from this requirement.

6.2.3 Variable Attenuator. A calibrated variable attenuator is needed to provide a means for adjusting the output indication level of the measuring instrument so as not to overload the frequency selective voltmeter or receiver. It may also be useful in checking for spurious responses.

6.2.4 Frequency-selective Voltmeter or Receiver. A frequency-selective voltmeter or receiver (may be more than one) which can tune to the carrier and any spurious output frequency of interest, is needed. If a receiver is used, it must have an output indicator. The receiver cabinet should incorporate good shielding techniques and the power lines should be well filtered.

6.2.5 Coaxial Switches. Coaxial switches (or suitable means for changing connections) may be required when using a calibrated signal generator for calibration purposes.

6.2.6 Measuring Equipment Enclosure. To prevent pickup of extraneous radiations during the measurements, the measuring equipment should (if necessary) be enclosed within a suitable shielded enclosure and the signal from the pickup device^{brought} into the shielded enclosure through a well-shielded cable.

6.2.7 Calibrated Signal Generator. A calibrated signal generator (or generators) to cover the carrier frequency and any spurious frequencies of interest is needed.

6.3 Measurement Procedure.

6.3.1 Standard Method. NOTE: The standard method ^{measures} ~~is used when~~ ~~it is desired to obtain~~ the absolute level of the radiated spurious ~~response~~ signal at the point of measurement without regard to the rated power output of the equipment under test. If the desired result is to be a

measurement of the absolute field intensity one must make the measurement in the "radiation field". Measurements made at lesser distances as may be required at low frequencies will involve the "induction field" and the results will require interpretation.

6.3.1.1 Connect the equipment as shown in Figure 6.

6.3.1.2 Operate the transmitter under test in its intended manner with its output connected to a ^{shielded} dummy load.

6.3.1.3 Tune the frequency-selective voltmeter or receiver through the frequency range of interest with the ^{calibrated} variable attenuator adjusted for maximum sensitivity of the measuring circuit. When a spurious output is found, orient the pickup device for maximum response and adjust the attenuator to obtain a suitable reference ^{indication} ~~reading~~ on the output indicator ^{of} ~~the~~ measuring instrument.

CAUTION: In all frequency-selective voltmeters or receivers spurious responses may occur by: (1) desensitization of the receiver by the entry of a strong off-channel signal through the antenna input. (2) ~~the~~ the entry of a strong on-channel signal through the receiver case or power lines, ^{such signals} ~~and~~ by-passing ^{the} ~~it~~ calibrated input attenuator. These responses must be known or determined for the particular device used.

In addition care must be taken to insure that the spurious signal being measured can be actually attributed to the equipment under test. This is easily determined by momentarily turning off the equipment under test.

6.3.1.4 If a frequency-selective voltmeter is used, calibrate it according to the recommended manufacturer's procedure and measure the spurious radiated signal. If a receiver is used it must be calibrated by means of appropriate calibrated signal generators.

6.3.1.5 The level of the spurious output is calculated as follows:

6.3.1.5.1 CW Spurious Output. CW Spurious output level (field intensity, DB above 1uV per meter) = meter reading in DB above 1uV (or substituted signal generator reading) + cable loss + antenna factor. *(between pickup device and the calibrated variable attenuator)*

6.3.1.5.2 Broadband Spurious Output. Spurious output level (field intensity, DB above 1uV per meter, per MC) = Meter reading (DB above 1uV) + cable loss (DB) + Antenna Factor (DB) - $10 \log_{10}$ (impulse bandwidth in MC).

6.3.1.6 To determine the radiation pattern and to evaluate path loss a multiplicity of measurement points varying in azimuth and radius should be used.

6.3.2 Alternate Method. NOTE: The alternate method measures an equivalent radiated power of the spurious signal.

6.3.2.1 Connect the equipment as shown in Figure 7. It should be noted that the pickup device must be located in the "radiation field" of the transmitter.

6.3.2.2 Operate the transmitter under test in its intended manner with its output connected to a shielded dummy load.

6.3.2.3 At the point of measurement tune the frequency-selective voltmeter or receiver through the frequency range of interest with the variable attenuator adjusted for maximum sensitivity of the measuring circuit. When a spurious output is found, *orient the pickup device for maximum response and* adjust the attenuator to obtain a suitable reference ~~value~~ indication on the output indicator of the measuring instrument.

~~6.3.2.4 The pickup device is oriented for maximum signal into the frequency selective voltmeter or receiver and an indication is obtained.~~

6.3.2.4 De-energize the transmitter.

6.3.2.5 A calibrated signal generator or other suitable source of

controlled oscillation is now used in conjunction with an appropriate radiating device to supply the substituted power. This radiating device is located at or in the near vicinity of the transmitter. Thus, essentially the same path is utilized for both the transmitter power and the substituted power.

6.3.2.6 Energize the calibrated signal generator and tune to the spurious frequency on the frequency selective voltmeter or receiver obtained in 6.3.2.3.

6.3.2.7 ^{Orient} ~~Rotate~~ the radiating device for maximum signal into the frequency-selective voltmeter or receiver.

6.3.2.8 Adjust the level of the calibrated signal generator to give the same indication at the frequency-selective voltmeter or receiver which was noted in 6.3.2.3.

6.3.2.9 To determine the substituted power, the input impedance and the effective gain of the radiating device must be known. The equivalent power of the spurious signal in watts can now be calculated as follows:

$$(1) \quad P = I^2 R_a$$

Where $R_a = R_g + (R + jX)$ *radiation resistance of the radiating device.*

~~* This to be completed at the next meeting of IRE 27th~~

I = current supplied by the signal generator.

$$(2) \quad I = \frac{E_g}{R_g + (R + jX)}$$

E_g = sig gen open-circuit voltage

R_g = signal generator internal impedance, (normally 50 ohms)

(R + jX) = Terminal impedance of the Radiating Device




Figure 6




FIGURE 7

MANEWA